



Reprocessed by ReNu Medical, Inc.

Instructions for Use Reprocessed DVT Garments

Reprocessed Device for Single Use

Symbol Legend:



Date of Reprocessing



Do Not Reuse



Consult instructions for use



Do not use if package is damaged



Non-Sterile – High Level
Disinfection



Caution See Instructions for use



ReNu Medical Catalog#



Customer ID#, if none specified;
ReNu Medical Catalog#



Reprocessor/Manufacturer



Not made with natural rubber Latex



Fragile, handle with care



Keep dry



Caution: Federal (USA) law restricts
this device to sale by or on the order
of a physician.



Original Equipment Manufacturer
Catalog#



ReNu Medical Sales Order



Qty of Devices included in Pkg/Cs



Reprocessed DVT Garments

Deep Vein Thrombosis Garments Description

Deep Vein Thrombosis (DVT) Garments are part of an external compression system, in which intermittent or sequential compression is provided using a pump/controller and limb garment. The system consists of the following three main components: a control unit, inflatable limb garments and conduit tubing with detachable connections.

Indications for Use

When coupled with an appropriate inflation system, garments are intended for use in preventing deep vein thrombosis (DVT), diminishing post-operative pain and swelling, enhancing blood circulation, and reducing wound healing time.

Additional Indications for Use

Clinical applications for the foot garments are prevention of venous stasis, assistance in the healing of cutaneous ulcers, including venous ulcers, reduction of acute & chronic edema, reduction of lower limb pain due to trauma or surgery, reduction of compartment pressures.

Contraindications

Reprocessed compression garments are contraindicated in the presence of the following conditions:

- Severe arteriosclerosis or other ischemic vascular disease.
- Congestive heart failure.
- Pulmonary edema.
- Known or suspected deep vein thrombosis or phlebitis
- Any local condition in which the garments would interfere.
 - Gangrene
 - Untreated infected wounds
 - Recent skin graft
 - Dermatitis

Note: If you are unsure of any contraindications refer to the patient's physician before using the device.

Warnings

- Prior to use, read and follow the Original Equipment Manufacturer's Operator's Manual for the compression pump pressure recommendations and pump compatibility.
- Do not repair or replace the tubing connectors as this may result in unwanted inflation of the garments.
- Do not operate in the presence of flammable gases (e.g. flammable anesthetics).
- Do not expose to excessive heat or freezing.
- Patients with diabetes, poor circulation, insensitive extremities, fragile skin, those on anticoagulation therapy and those predisposed to tissue viability problems should receive special attention. Use the lowest effective pressure and timing and additional padding.
- Check the patient every 8 to 12 hours for skin reddening and any early signs of tissue viability problems.
- Clinical judgement is required to determine if the patient's skin condition requires additional protective measures, or if the therapy should be discontinued.
- Compression therapy may contribute to circulatory failure if excess inflation pressure is applied or if patient has peripheral vascular ischemic disease.

ReNu Medical, Inc.
830 80th Street SW Suite 100
Everett, WA 98203
www.renumedical.com
877-252-1110



Reprocessed DVT Garments

- Compression therapy may increase the risk for compartment syndrome or peripheral neuropathy.

Precautions

- Proper garment sizing and application is essential.
- Ensure proper garment positioning to lower the risk of pressure points on the limb.
- Use of anti-embolism stockings under compression devices may provide greater comfort to the patient.
- Ensure proper connections to the external pump controller.
- Ensure that tubing is not kinked or twisted as this could restrict airflow.
- Maximum inflation pressure should not exceed patient's diastolic pressure. Check for dorsalis pedis and posterior tibial pulses during maximal inflation.
- Do not elevate patient's feet above the level of the heart.
- Immediately remove garments if patient experiences numbness, tingling or leg pain.
- To minimize local air movement, turn garment cooling off when using garments in the operating room.
- If compression is interrupted for more than 30 minutes in patients at risk for deep venous complication, resume continuation of the compression therapy only after noninvasive reevaluation of the new situation.
- Lower limb position in relation to the garment and tubing should also be considered particularly when a patient is unconscious or has reduced sensation or the ability to move their legs.
- Additional care should be taken when placing the garments on any deformed leg or foot, or on legs with significant edema.

Additional Precautions

Garments Designed for Foot Application:

Refrain from walking and weight bearing while wearing foot garments.

Periodically check the skin condition under patients wearing stockings or stockinette's.

Directions for Use

1. Remove the device from the package.
2. Inspect the device for overall condition and physical integrity. Do not use the device if any damage is noted.
3. Prepare garment for correct positioning following the directions and symbols/icons on the inside of the garment or otherwise provided by the original equipment manufacturer.
4. Center patient's limb on the inside of the garment.
5. Before starting to wrap the garment around the leg, make sure the pump controller is in the OFF position.
6. Start by wrapping the side without a hook tape or fastener. Follow by wrapping the side with fasteners.
7. Wrap the garment snugly around the patient's limb with the inflatable bladder on the rear side of patient's extremities. Attach the hook/fastener securely to the garment, starting with the ankle of the patient's limb. Achieve a snug and secure, but not too tight, fit around all sections of the patient's limb.
8. If more than one limb is to receive treatment, repeat the above steps on the other side.

ReNu Medical, Inc.
830 80th Street SW Suite 100
Everett, WA 98203
www.renumedical.com
877-252-1110



Reprocessed DVT Garments

9. In the case of single leg application, refer to the original equipment manufacturer Operation Manual for Compression Therapy Settings.
10. Do not position the garment such that the tubing can form pressure points on the patient's limb. If a patient will be placed in certain surgical positions like kneeling or similar positions, rotation of the garment with the tubing facing away from the patient will prevent pressure points.
11. Before attaching the garment to the air tubing, make sure the tubing is not kinked or twisted.
12. Attach the air tubing to the pump. Push the connectors together firmly to properly engage. To uncouple the connectors, firmly pull them apart.
13. Adjust the pump pressure to the recommended pressure setting for the garment in use, unless otherwise directed by the physician.
14. Turn the control unit ON after the tubing is correctly attached to the garment and the control unit.
15. Depress the garment-cooling button, if cooling is desired.
16. Device is intended for use during a single patient procedure.
17. Refer to original equipment manufacturer literature for pump compatibility, accessory information and further guidance

Additional Directions for Use

Garments Designed for Foot Application

1. Place the inflatable bladder under the arch of foot.
2. Apply stocking or stockinette over foot and ankle and smooth any wrinkles.
3. Close the fastener over the top of the foot. Next bring the rear strap around the heel and close the fastener.
4. Position the foot below heart level during pump operation for best results.
5. Inflate only after proper placement.

The names of the actual Original Equipment Manufacturer and products mentioned within this document and any information listed on the label is provided as identification prior to High-Level Disinfection reprocessing and may contain trademarks of unrelated third parties who may not represent the device after reprocessing.

ReNu Medical, Inc.
830 80th Street SW Suite 100
Everett, WA 98203
www.renumedical.com
877-252-1110